


DESCRIPTION	DATE
Installed Strip Heater	Feb. 1/06

Test: Temperature Profile: after 9 minutes

## INFRARED IMAGE



## PHOTO



IR information	Value
Date of creation	01/02/2006
Time of creation	10:28:40 AM
Object parameter	Value
Ambient temperature	21.0°C
Label	Value
SP01	24.7°C
SP02	22.8°C

## DESCRIPTION:

Infrared image of an installed Kendrick Astro Instruments strip heater.

Time frame: after approximately 9 minutes of operation time.

Temperature profile: strip heater to telescope optics temperature variation: 1.9 degrees C.